



Bid Number/बोली क्रमांक (बिड संख्या):

GEM/2025/B/5805194

Dated/दिनांक : 15-01-2025

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	14-02-2025 14:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	14-02-2025 14:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	180 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Education
Department Name/विभाग का नाम	Department Of Higher Education
Organisation Name/संगठन का नाम	Delhi University
Office Name/कार्यालय का नाम	Vallabhbhai Patel Chest Institute
क्रेता ईमेल/Buyer Email	buyer21.du.dl@gembuyer.in
Total Quantity/कुल मात्रा	1
Item Category/मद केटेगरी	CT Scanners 64 Slice (Q2)
MSE Exemption for Years Of Experience/अनुभव के वर्षों से एमएसई छूट/ and Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है	Yes
Startup Exemption for Years Of Experience/अनुभव के वर्षों से स्टार्टअप छूट/ and Turnover/ टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है	Yes
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Certificate (Requested in ATC), Additional Doc 1 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Do you want to show documents uploaded by bidders to all bidders participated in bid?/	No
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No
Type of Bid/बिड का प्रकार	Two Packet Bid
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	3 Days

Bid Details/बिड विवरण	
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Estimated Bid Value/अनुमानित बिड मूल्य	55000000
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation
Arbitration Clause	No
Mediation Clause	No

EMD Detail/ईएमडी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
EMD Amount/ईएमडी राशि	1375000

ePBG Detail/ईपीबीजी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
ePBG Percentage(%) / ईपीबीजी प्रतिशत (%)	5.00
Duration of ePBG required (Months) / ईपीबीजी की अपेक्षित अवधि (महीने).	62

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई कैटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

Beneficiary/लाभार्थी :

Director
Deputy Registrar Office, Administrative Block, 3rd Floor, M.S. Building, VPCI, University of Delhi, Delhi-110007
(The Director, Vpci)

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता	Yes
--	-----

MSE Purchase Preference/एमएसई खरीद वरीयता

--	--

MSE Purchase Preference/एमएसई खरीद वरीयता

Yes

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.

2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to their meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.

3. Preference to Make In India products (For bids < 200 Crore): Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate. The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

OM No.1 4 2021 PPD dated 18.05.2023 for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

4. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 OM No.1 4 2021 PPD dated 18.05.2023 for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

5. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

Pre Bid Detail(s)

Pre-Bid Date and Time	Pre-Bid Venue
30-01-2025 15:00:00	Committee Room, 3rd Floor, Multistorey Building, VPCI

CT Scanners 64 Slice (1 pieces)

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

Technical Specifications/तकनीकी विशिष्टियाँ

* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
TYPE	Number of slices per rotation	64
	Number of Detector Rows	64
	Number of acquisition channels	64
DETECTOR ASSEMBLY	Extended Field of View (cm)	50
	Total detector width, z-axis	40.0 (millimeter)
DETECTOR PERFORMANCE	High-contrast spatial resolution at 0% MTF with full FOV	≥ 17 lp/cm
ADVANCED IMAGE ACQUISITION	Cardiac	No
	Extended coverage for 4-D imaging, Max coverage with 1, 1 decimal 5 and 3 sec sampling resolution	No
	Contrast Enhanced Dual-energy acquisition	Yes, No
GANTRY	Minimum Gantry aperture	70.0 (centimeter)
X-RAY GENERATOR	Minimum kW output	32, 55, 72, 80 Or higher (kiloWatt)
	kVp range	80- 140
X-RAY TUBE	Minimum Anode Heat storage (MHU)	3.5
	Anode Heat dissipation rate	≥ 780 kHU/min
	Tube cooling	Oil
	Max mA	≥ 400 mA
	Max scan time (sec)	≥ 100 sec
PATIENT TABLE	Table Type	Moveable
	Range of movement - Vertical (cm)	40 - 100 cm or better
RADIATION DOSE	Pediatric-specific dose control	Yes
	Prospective ECG gating	No
	Retrospective ECG editing	No

Specification	Specification Name/विशिष्ट क नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Iterative image reconstruction for dose reduction	Yes
	Overbeaming reduction sliding collimation	Yes
	Low dose Cardiac axial acquisition	No
CLINICAL APPLICATIONS AND FUNCTIONALITY	Auto vessel mapping	Yes
	Left Ventricular output	No
	Myocardial evaluation	No
	Lung nodule assisted reading	Yes
	Lung nodule CAD	Yes, No
	Lung function analysis	Yes
	Respiratory gating	No
	calcium scoring	No
	Virtual colonoscopy CAD	No
	Vessel analysis (noncardiac)	Yes
	Brain perfusion	Yes
	Z-axis coverage for brain perfusion	Yes
	Auto bone removal	Yes
	Body perfusion	No
	Having Highest achievable temporal resolution	No
	Dental CT	Yes
	CT Fluoro / interventions with hardware and software	No
	Neuro DSA	Yes
	Metal artifacts reduction	Yes
IMAGE RECONSTRUCTION	Max reconstruction rate, (512 x 512), image/sec	20
SYSTEM INTEGRATION	DICOM Worklist License	Yes
IMAGE PROCESSING	OEM postprocessing workstation with Single/Dual medical grade Monitor	Standalone

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	If it is server workstation, number of nodes to be provided	4, Not Applicable in case of standalone workstation
	Number of concurrent licenses to be provided	Not Applicable in case of standalone workstation
	Minimum Rendering Capacity of Server	Not Applicable in case of standalone workstation
SERVER HARDWARE	Minimum Image storage capacity of Server	2, 4 Or higher (terabyte)
	Minimum Ram Size of server	0, 32, 64 Or higher (gigabyte)
	Size of Monitor of server	19 inch or more
	Type of Monitor of server	LCD/LED/TFT
WORKSTATION HARDWARE	Minimum Ram Size of Workstation	8, 16, 4 Or higher (gigabyte)
	Minimum Storage capacity of Workstation	500, 1000 Or higher (gigabyte)
WORKSTATION APPLICATIONS	Neuro DSA for workstation application	Yes
	Cardiac Angio	No
	Cardiac calcium scoring	No
	CT angio	Yes
	Filming	Yes
	CT VRT Imaging	Yes
	Perfusion imaging in workstaion	Yes
CONTRAST BASED DUAL ENERGY APPLICATIONS	Virtual Non-contrast imaging	No
	Mono-energetic imaging	No
	Direct Angio	No
	Lung Perfusion	No
	Gout Imaging	No
	Calculi Characterization	No
	Bone Marrow imaging	No
	Iodine mapping	No
ACCESSORIES	Online UPS for full backup of atleast 30 minutes	Yes
	Lead aprons - 5 Nos with Hanger to be provided	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Gonad and thyroid Shield - 5 No each	Yes
	Pressure Injector	Single head, Dual head
	Color Printer	Yes
	Head Holder	Yes
	Lead Glass Setup to be provided	Yes
	X-Ray Film viewer	Yes
	Variety of phantoms for quality control	Yes, No
CERTIFICATIONS AND REPORTS	Availability of test report/QA and QC report during warranty and CMC as per AERB/NABH/JCI guidelines	Yes
	Confirmity to Manufacturer's Certification	ISO 9001 & ISO 13485
PRE-INSTALLATION SITE REQUIREMENT	Preparation of CT Scan Room , console Room , Technical room, one wash room etc to be done by the seller in the area of approx 800 sq feet	Yes
	Tile based false ceiling and floor tiling and wall tiling till false ceiling in rooms to be done by the seller	Yes
	Necessary modification civil work, electrical , plumbing, furnishing, air conditioning in the given area to be done by the seller	Yes
	Doors other than main CT room to be of good quality Aluminium doors and should be done by the seller	Yes
	Any other necessary work for system functioning like earthing , proper LED lighting power points etc to be done by the seller	Yes
	Room shielding to be done by the seller	As Per AERB Guideline
INSTALLATION & TRAINING	Number of installations of the CT in Central /State/PSU Govt Hospitals(Note: Seller should supply a performance certificate of the device to the buyer)	upto 3

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
WARRANTY	Warranty in Years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	5 Or higher

Additional Specification Parameters - CT Scanners 64 Slice (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
(1) MAIN OPERATOR CONSOLE- The main console should have standard functions like	(1) Pulmonary Angiography, HRCT Lung, 3D VRT, 3D SSD, MPR, MIP, CT Angiography, Virtual bronchoscopy & endoscopy, Vascular assessment, 3D artefact suppression, Auto bone removal, Metal artefact reduction, etc. Emphysema Quantification and lung nodule evaluation Software should be provided. There should be direct generation of axial, sagittal, coronal or double-oblique images from standard scanning protocols.
(2) MAIN OPERATOR CONSOLE- The main console should	(2) * support archiving facility on DVD&CD. DICOM 3.0 facility to send, store, print, receive, query/retrieve etc. should be standard. * should be capable of integrating with departmental RIS and HIS.
(3) Image Processing Software (4) ACCESSORIES: Multi tray Dry/Laser Imager of reputed make with minimum 600 dpi and compatible with the CT System - One	(3) Review workstation - MPR, MIP, 3D VRT formats, SSD, Virtual Bronchoscopy and endoscopy, tools for stenosis measurement, Advanced vessel analysis, Automatic bone removal etc. Emphysema Quantification and lung nodule evaluation Software. Archival on DVD/CD/USB and DICOM print facility. (4) Yes
(5) Lung Nodule CAD (6) Reconstruction Matrices (7) Product Certification (8) Pressure Injector (9) Minimum RAM size of Workstation (Gby) (10) Minimum Storage Capacity of workstation (Gby)	(5) Yes (6) 1024x1024 (7) CDSCO/EU-CE (8) Dual Head (9) 32 Gigabyte (10) 2 Terabyte
Please See Complete Tender Document as per e-tender notice no. VPCI/PUR/OT/CAP/GEM/24-25/19-5/133 (PDF File) attached in Buyer Added Bid Specific Terms & Conditions attached in Point no. 2.	All Bidders are requested to go through the Tender document carefully and submit all the required Annexures with the Technical Bid while submitting the bid

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/प्रेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेक्षित/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Sunil Kumar	110007,VALLABHBHAI PATEL CHEST INSTITUTE, NORTH CAMPUS, OPPOSITE SHRI RAM COLLEGE OF COMMERCE, UNIVERSITY OF DELHI	1	30

Special terms and conditions-Version:3 effective from 21-10-2022 for category CT Scanners 64 Slice

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules made there under as amended till date will always be applicable. This will include all notifications issued by *Central Drugs Standard Control Organisation(CDSCO)*, Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
3. The price offered by the seller shall not, in any case exceed the DPCO controlled price, if any, fixed by the Central/State Government, the Maximum Retail Price (MRP) and the selling price. The seller must reduce the prices if there is any reduction in DPCO ceiling price, if any.
4. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
5. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address,telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled.Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
6. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
7. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
8. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.
OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its

- machines/equipments supplied.
9. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
 10. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent. In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
 11. **Software:** All software updates should be provided free of cost during warranty period.
 12. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

2. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

Technical Bid

1. Please See Complete Tender Document as per e-tender notice no. VPCI/PUR/OT/CAP/GEM/24-25/19-5/13 (PDF File) attached in Buyer Added Bid Specific Terms & Conditions attached in Point no. 2.
2. All Bidders are requested to go through the Tender document carefully and submit all the required Annexures with the Technical Bid while submitting the bid
3. The demonstration of the quoted model of CT Scan Machine is necessary during the technical Evaluation of bid/s.
4. The bidder should assure compliance of the quoted model of the CT Scan Machine with the attached completed specifications of the required CT Scan Machine

5. Pre-bid Meeting:-

The pre-bid meeting shall be held in the Committee Room, 3rd Floor, Multistorey Building, VPCI on the scheduled date & time (30.01.2025 at 03:00 p.m.). Non-attendance of pre-bid meeting will not be a cause for disqualification of the bidder and it shall be presumed that the bidder does not require any clarification. The purpose of Pre-bid meeting is to clarify the issues raised by bidders. The queries/clarifications shall be submitted in the format as provided in Annexure-XV of the tender. Each Bidder shall send only one e-mail containing the Pre-Bid Queries, if required. In case of more than one e-mail only the queries provided in the first e-mail shall be considered. VPCI reserves the right to change the time and venue of the Pre-Bid Meeting.

3. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for attached categories, trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.

This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद

पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।

---Thank You/धन्यवाद---