



बिड संख्या/Bid Number: GEM/2025/B/6612330

दिनांक /Dated: 13-10-2025

बिड दस्तावेज़ / Bid Document

बिड विवरण/Bid Details			
बिड बंद होने की तारीख/समय /Bid End Date/Time	04-11-2025 19:00:00		
बिड खुलने की तारीख/समय /Bid Opening Date/Time	04-11-2025 19:30:00		
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)		
मंत्रालय/राज्य का नाम/Ministry/State Name	Ministry Of Education		
विभाग का नाम/Department Name	Department Of Higher Education		
संगठन का नाम/Organisation Name	Delhi University		
कार्यालय का नाम/Office Name	Vallabhbhai Patel Chest Institute		
कुल मात्रा/Total Quantity	1		
वस्तु श्रेणी /Item Category	C Arm Fluoroscope X - Ray Machine (V2) (Q2)		
टर्नओवर के लिए एमएसई को छूट प्राप्त है / MSE Exemption for Turnover	Yes Complete		
टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है / Startup Exemption for Turnover	Yes Complete		
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	Certificate (Requested in ATC),Additional Doc 1 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer		
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेज़ों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)		
बिड लगाने की समय सीमा स्वतः नहीं बद्धने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	1		
देनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बद्धई जाएगी। / Number of days for which Bid would be auto-extended	7		

बिड विवरण/Bid Details			
ओंटो एक्सर्टेशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	3		
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	No		
बिड का प्रकार/Type of Bid	Two Packet Bid		
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	5 Days		
निरीक्षण आवश्यक (सूचीबद्घ निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No		
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation		
वित्तीय दस्तावेज की आवश्यकता है / Financial Document Required	Yes		
मध्यस्थता खंड/Arbitration Clause	No		
सुलह खंड/Mediation Clause	No		

ईएमडी विवरण/EMD Detail

एडवाईजरी बैंक/Advisory Bank	State Bank of India	
ईएमडी राशि/EMD Amount	120000	

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईपीबीजी प्रतिशत (%)/ePBG Percentage(%)	5.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	62

- (a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.
- (b).ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance securityshould be in favour of Beneficiary, wherever it is applicable.

लाभार्थी /Beneficiary :

Director

Deputy Registrar Office, Administrative Block, 3rd Floor, M.S. Building, VPCI, University of Delhi, Delhi-110007

(Director, Vpci)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता / MII Purchase Preference

एमआईआई खरीद वरीयता / MII Purchase Preference	Yes
/ Purchase Preference to MII sellers availabele upto price within L1+X%	20
Maximum Percentage of Bid quantity for MII purchase preference	50
/ Allow participation only from Class 1/Class 2 local suppliers as per the Public procurement(Preference to Make-in-india) order 2017 date 16.09.2020(as amended and applicable time to time)	Yes, in compliance with the MII ORDER : DPIIT Order(as amended and applicable time to time)

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
/ Purchase Preference to MSE OEMs available upto price within L1+X%	15
/ Maximum Percentage of Bid quantity for MII purchase preference	25

1. If the bidder is a Micro or Small Enterprise (MSE) as per latest orders issued by Ministry of MSME, the bidder shall be exempted from the eligibility criteria of "Bidder Turnover" as defined above subject to meeting of quality and technical specifications. If the bidder itself is MSE OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking exemption from Turnover, shall upload the supporting documents to prove his eligibility for exemption.

2. If the bidder is a DPIIT registered Startup, the bidder shall be exempted from the the eligibility criteria of "Bidder Turnover" as defined above subject to their meeting of quality and technical specifications. If the bidder is DPIIT Registered OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking exemption from Turnover shall upload the supporting documents to prove his eligibility for exemption.

3. Preference to Make In India products (For bids < 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate. The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

OM No.1 4 2021 PPD dated 18.05.2023 for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

4. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and

its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 OM No.1 4 2021 PPD dated 18.05.2023 for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

5. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

एक्सेल में अपलोड किए जाने की आवश्यकता /Excel Upload Required:

financial bid - 1760356940.xlsx

C Arm Fluoroscope X - Ray Machine (V2) (1 pieces)

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification विशिष्टि का नाम /Specification Name		बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)	
PRODUCT Type of Mechanical Motion of C Arm		Manual	
	Angulation Movement of C Arm (in degrees)	± 200	
	Range of Frequency of X ray Generator in khz	More than 100	
	Generator Power output in Kw	3.5	
	Anode heat storage capacity with tolerance ±10%	50 KHU	
	Number of Monitors provided on/with intensifier system	Single monitor with dual image display touch screen	
	Camera Resolution in pixels	1K x 1K	

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
WARRANTY	Warranty in Years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	1, 2, 3, 4, 5 (year)

Additional Specification Parameters - C Arm Fluoroscope X - Ray Machine (V2) (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)	
Terms & Conditions related to required technical specifications	1. The Specification provided in GeM tender document is only for reference and for actual specifications bidders are requested to refer the tender document as ATC under the heading "Buyer Added Bid Specific Terms & Conditions" 2. The technical evaluation will be as per these specifications only, and the bidder shall quote the equipment and submit the technical compliance accordingly.	

^{*} Bidders offering must also comply with the additional specification parameters mentioned above.

परेषिती /रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Sunil Kumar	110007,VALLABHBHAI PATEL CHEST INSTITUTE, NORTH CAMPUS, OPPOSITE SHRI RAM COLLEGE OF COMMERCE, UNIVERSITY OF DELHI	1	60

Special terms and conditions-Version:1 effective from 10-07-2024 for category C Arm Fluoroscope X - Ray Machine (V2)

- All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017)
 made there under as amended till date will always be applicable. This will include all notifications
 issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare
 (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time
 in this regard.
 - The sellers are registered on GeM based on the self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
 - 3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
 - 4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.

- 5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
- 6. Comprehensive warranty: Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
- 7. Service centres: Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
- 8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
- 9. Packing and Marking: Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transhipment (if any), rough handling, open storage etc. without any damage, deterioration etc. The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
- 10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
- 11. Installation, Training, Manuals: Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
- 12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent. In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
- 13. Software: All software updates should be provided free of cost during warranty period.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्ते/Buyer Added Bid Specific Terms and Conditions

SN	Specification Name	Values
	Performance Parameters	
1.	Type of Mechanical Motion of C Arm	Manual
2.	Angulation Movement of C Arm in degrees	± 200
3.	Source to Image Distance in mm	951-1000
4.	Arc Orbital Movement of C Arm	At least +90° - 40°
5.	Panning Movement of C Arm (Swivel movement)	±12.5º
6.	Range of Frequency of X ray Generator in khz	More than 100
7.	Generator Power output in Kw	3.5
8.	Fluoroscopy Range in kvp	40-110
9.	Fluoroscopy mA range, maximum in mA	0.2 to 7
10.	Dual focus Rotating Anode X-Ray Tube of focal spot	0.6mm² (small) & 1.5m m² (large)
11.	Anode heat storage capacity with tolerance ±10%	50 KHU or more
12.	Modes available	Continuous fluoro, Pulse d Fluoro, Radiographic mode
13.	Pulses per sec	1 to 15
14.	Mode selection for Kv and mA	Automatic
15.	Type of Image Intensifier	9 " Dual Field
16.	Number of Monitors provided on intensifier system	single monitor with dual image display touch scr een
17.	Type of monitor to be provided with intensifier system	LCD/LED/TFT
L8.	Size of Monitors in inch	29-32

19.	Type of Collimator controlled	Both Fixed Collimation Slot & Motorized IRIS
20.	Type of Camera	CCD Camera
21.	Camera Resolution in pixels	1K × 1K
22.	Acquisition of Camera	16 bits
23.	Image storage capacity in images	20000 to 50000
24.	Zoom function available	2x
25.	DICOM Compatibility	Yes
26.	The system should be ready to connect with HIS/ PACS	Yes
27.	Data transfer	USB
28.	Radiography through Cassette with holder	No
29.	Size of Cassette holder provided (for radiographic mode) in in ch x inch	Not provided
30.	Upgradable option to digital substration of angigraphy and peripheral angiography	Yes
31.	UPS Availability	External
32.	UPS backup time in minutes	atleast 15 min
33.	Servo Stabilizer supplied	Yes
34.	Rating of Servo Stabilizer	Not provided
35.	Lead apron to be provided Lead apron should be lead free light weighted (BARC certified) Thyroid shield Gonad shield Lead goggles	5 5 5 5
36.	Digital C-Arm with flat panel detection and console screen	flat panel detection wit h size 23x23 or more, 15" or more LCD touch s creen. Control panel wit h live display.

37.	Anode heat storage capacity	Data sheet of tube shou ld be submitted Radiographic / digital sp ot mA - 70 mA or more. Generator should be 22 0 Khz or more	
38.	Radiographic / digital spot		
39.	Warranty (Option of comprehensive warranty is available thro ugh bidding only, which if opted will supersede normal warranty in the catalogue)		
40.	Demonstration, if required has to be given		
41.	Five years of warranty		
42.	List of consumables to be given along with rat es which will be freezed for five years		
43.	The equipment should (a) be conform to the s tandards laid down by the Bureau of Indian St andards (BIS) and if not then device shall (b) c onform to the standard laid down by the International Organization for Standardization (ISO) or the International Electro Technical Commi ssion (IEC), or by any other pharmacopoeial st andards and if not (a) or (b) then (c) the device shall conform to the validated manufacturer's standards		
44.	Machine should have ICMED certifications		
45.	The system should be Microprocessor Controll ed, Compact and Mobile Offering Excellent Im age Details	Yes	
46.	Radius (Immersion Depth) of C Arm in Mm : 61 0-750	610-750	
47.	Image Reversal	Yes	
48.	Monitor with Trolley	Yes	
49.	Sterile Cover For C-Arm::X-Ray Tube::Image I ntensifier	Yes	

50.	Radiation::Fuoro Exposure Deliver Switch	Hand Switch and Foot Swit
51.	Display of MA, KV and other safety feature on C-Arm	Yes
52.	Patient Demographic Detail Entry Facility	Yes
53.	Type of Power Supply	Single phase, 220-240 Vac, 50HZ-16 Amps Socket
54.	Compliance to Medical Device Rules (MDR) 20 17 as amended till date	Yes
55.	Availability of Valid Medical Device License for the Product Issued from the Competent Autho rity defined under Drugs and Cosmetic Act 19 40 and rules made there under as amended til I date	Yes
56.	Horizontal Movement of C Arm (in Mm)	170-240
57.	Vertical Movement of C Arm (in Mm)	400-450
58.	Safety Feature Available for C Arm Movement	Locks for all the movement s, control stand foot lock, S teering Wheel for easy Ste ering & movement.
59.	Automatic Image Parameter selection with ca pability or switching on to Manual selection	Yes
60.	Submission of all necessary certifications, Lice nse and test reports to the buyer at the time o f bid submission or along with supplies as per buyer requirement	Yes
61.	AERB Type approved equipment	Yes
62.	Conformity to Indian Standard IS 7620-Part 1, Part 2, Part 3	Yes

Additional Specification Parame	ters - C ARM FLUOROSCOPE X RAY MACHINE
Specification Parameter Name	Bid Requirement (Allowed Values)

Digital C Arm with FPD & Certifiation	Should have Dynamic Flat panel detector IGZO technology for low dose imaging with detector size 9inch \times 9inch or m ore , Pixel pitch of $150\mu m$ or less, Pixel Matrix : $1.5K \times 1.5K$ or more and spatial resolution of $3.0lp/mm$ or more, Certificat ion : ISO/ICMED/BIS
Control Panel	Dual LCD display for selection of exposure parameters, & Display for all operative parameter. Should have Automatic Dose Rate Control Mode with KV rise time of 40KV/sec or more.
" C" Movement	Depth of "C" should be 680mm or more , Free Space of "C" Should be 800 mm or more. Total Width (Foot print) should be 800mm or less. Double Steering wheel system for easy mo vement and positioning of machine.
Fluoroscopy modes	Control Panel should have provision to selects following exposure modes; Continuous Flouro: 0.2 to 4.5mA in normal mode & 0.6 to 8 mA in Boosted fluoroscopy Mode, Puls ed Fluoroscopy: 0.6 mA to 8 mA and Low dose or Pae diatric Fluoro: 0.1 mA to 2.25 mA or less.
Frequency & Radiation Protection Accessories	High frequency X ray generator having frequency of 220 KHz or more All standard accessories as per product listing on GEM portal should be supplied with machine i.e. Lead Aprons - Full bod y) 5 number each, No of Thyroid shield and Gonald Shield - 5 Number each. Lead Aprons stand with 8 hangers - 1 No., Lead Googles - 05 Nos

Note: All bidders are requested to submit Technical bid and financial bid as per the E-Procurement Tender Notice no.VPCI/PUR/OT25-26/Capital/3-51 (PDF File) attached in Buyer Added Bid specific Terms & Conditio ns.

2. Buyer Added Bid Specific ATC

Buyer uploaded ATC document Click here to view the file.

अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

- 1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
- 2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to

- exemption provided to such sellers under GeM GTC.
- 3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
- 4. Creating BoQ bid for single item.
- 5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
- 6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
- 7. Floating / creation of work contracts as Custom Bids in Services.
- 8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for attached categories, trials are allowed as per approved procurement policy of the buyer nodal Ministries)
- 9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
- 10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
- 11. Creating bid for items from irrelevant categories.
- 12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
- 13. Reference of conditions published on any external site or reference to external documents/clauses.
- 14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
- 15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
- 16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.

यह बिंड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of Gem GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in

---धन्यवाद/Thank You---